



Qube Medical Products Sdn. Bhd.

No. 9 & 10, Jalan KPK 1/3, Kawasan Perindustrian
Kundang, 48020 Kundang, Selangor, Malaysia.
Tel.: +603 6034 5857 Fax.: +603 6034 5859

SPECIFICATION FORM INFORMATION

Glove Name/ Description : Powder- Free Nitrile Examination Gloves Regular Blue (Non-Sterile)
Factory Name : Qube Medical Products Sdn Bhd
Factory Location : Kundang Jaya
Date Completed : 17.09.2016
Revision : 01

PHYSICAL SPECIFICATIONS

Colour : Regular Blue **Cuff-Style** : Beaded
Product Weight : 6.0 g +/- 0.2 **Palm Pattern** : Smooth
Material Type : Nitrile Rubber **Donning Surface Finishing** : Chlorinated
Antimicrobial : Biostat G8 **Orientation** : Ambidexterous

GLOVE SPECIFICATIONS

	Extra Small	Small	Medium	Large	Extra Large	Variance
Overall Length (mm)	290	290	290	290	290	+/- 5 mm
Weight of Glove (g)	5.0	5.5	6.0	6.5	7.0	+/- 0.2 g
Cuff Thickness (mm)	0.070	0.070	0.070	0.070	0.070	+/- 0.02 mm
Palm Thickness (mm)	0.090	0.090	0.090	0.090	0.090	+/- 0.02 mm
Finger Thickness (mm)	0.130	0.130	0.130	0.130	0.130	+/- 0.02 mm
Cuff Width (mm)	80	84	90	95	97	+/- 5 mm
Palm Width (mm)	75	83	94	105	115	+/- 5 mm
Wrist Width (mm)	73	79	85	92	97	+/- 5 mm
Thumb Length / Width (mm)	50 / 25	55 / 32	60 / 32	63 / 38	65 / 40	+/- 2 mm
Finger 2 Length / Width (mm)	65 / 23	70 / 26	75 / 30	75 / 36	77 / 38	+/- 2 mm
Finger 3 Length / Width (mm)	75 / 25	78 / 28	82 / 32	85 / 40	86 / 44	+/- 2 mm
Finger 4 Length / Width (mm)	77 / 24	80 / 26	85 / 30	87 / 34	90 / 36	+/- 2 mm
Finger 5 Length / Width (mm)	50 / 20	55 / 22	58 / 25	63 / 29	65 / 32	+/- 2 mm

PERFORMANCE SPECIFICATIONS

AQL (Average Quality Level) : G1, 2.5 **Tensile at Break / Before Aging** : min 14 Mpa
Pin-Hole : G1, 2.5 **Tensile at Break / After Aging** : min 14 Mpa
Visual Major : G1, 2.5 **Elongation at Break / Before Aging** : 500 %
Visual Minor : G1, 4.0 **Elongation at Break / After Aging** : 400 %
Quality Dimensions : S2, 4.0 **Quality Management Certifications** : ISO 9001:2008 &
ISO 13485:2003

Prepared By:

QA Executive

Date: 17.09.2016

Date : 17.09.2016